

Alarm Status

Low		OK
High		Fail

Recorder Info

Serial # : 0010017494	Model : HAXO-8	Battery : OK	Trip # : 115
User ID : cmro II STATIE TROACA 09 06			

Recorder Configuration

Start type : Date/Time start	Temperature alarms	Humidity alarms
Start delay : None	Lower : -30.0 °C after 2 Consecutive	Lower : 10.0 %RH
Interval : 5 Minutes	Upper : 50.0 °C after 2 Consecutive	Upper : 100.0 %RH
Alert indicator : Enabled lower & upper		
OK indicator : Enabled		

Recorded Data


First reading : 12/16/2016 5:20:00 PM	Temperature statistics	Humidity statistics
Last reading : 12/18/2016 7:50:00 AM	Lowest : -8.5 °C	Lowest : 74.6 %RH
Elapsed Time : 1 Day, 14 Hours, 30 Minutes	@ 12/17/2016 7:30:00 AM	@ 12/17/2016 1:15:00 PM
Total readings : 463	Highest : 3.8 °C	Highest : 100.0 %RH
First evaluated : 12/16/2016 5:20:00 PM	@ 12/17/2016 12:30:00 PM	@ 12/17/2016 12:30:00 PM
Last evaluated : 12/18/2016 7:50:00 AM	Average reading : -2.9 °C	Average reading : 80.7 %RH
Evaluated Time : 1 Day, 14 Hours, 30 Minutes	Standard Deviation (S) : 3.3 °C	Standard Deviation (S) : 2.9 %RH
Evaluated Readings : 463	Mean Kinetic : -2.20 °C	

Low Alarm

Temperature	Triggered : (none)
	Time below : (none)
	Occurrences : 0
	°C - Minutes below : 0.00
Humidity	Triggered : (none)
	Time below : (none)
	Occurrences : 0


High Alarm

Temperature	Triggered : (none)
	Time above : (none)
	Occurrences : 0
	°C - Minutes above : 0.00
Humidity	Triggered : @ 12/17/2016 12:30:00 PM
	Time above : (none)
	Occurrences : 1



Chart

0010017494 - cmro II STATIE TROACA 09 06



UTC +02:00, standard time

Index	Date	Time	%RH	°C
463	12/18/2016	7:50:00 AM	83.9	-3.6

Summary

LogTag ID	0010017494	
User ID	cmro II STATIE TROACA 09 06	
LogTag battery	OK	
Non alert range	10.0 to 100.0 %RH	-30.0 to 50.0 °C
Time zone	UTC +02:00, standard time	
Number of readings	463	
Reading interval	5 Minutes	
Number of starts	115	
First reading	12/16/2016 5:20:00 PM	
Last reading	12/18/2016 7:50:00 AM	
Elapsed Time	1 Day, 14 Hours, 30 Minutes	
Reading range	74.6 to 100.0 %RH	-8.5 to 3.8 °C
Average reading	80.7 %RH	-2.9 °C
Standard Deviation (S)	2.9 %RH	3.3 °C
Degree Minutes below lower alert	0.00 %RH-Minutes	0.00 °C-Minutes
Degree Minutes above upper alert	0.00 %RH-Minutes	0.00 °C-Minutes
Mean Kinetic Temperature	-2.20 °C	
Time below lower alert	None	None
Time above upper alert	None	None
Time not in alert	1 Day, 14 Hours, 30 Minutes	1 Day, 14 Hours, 30 Minutes